

CORRIGENDA**Corrigendum to Commission Decision 2009/886/EC of 27 November 2009 amending Decision 2002/364/EC on common technical specifications for in vitro diagnostic medical devices**

(Official Journal of the European Union L 318 of 4 December 2009)

On page 37, in the Annex, Table 6, in the first row, last column:

for: '≥ 95 % agreement between genotyping and sequencing',

read: '> 95 % agreement between genotyping and sequencing'.
